

Item	Section 1 - Personnel	Reference	Yes-No or NA	
1.1	Does the analyst(s) interviewed meet the job description position requirements, training and qualifications for performing the test? Supervisor: _____ Technician: _____	(G25)6.1		

Item	Section 2 - Equipment & Facilities	Reference	Yes-No or NA	
2.1	Are the BOD bottles water sealed, with a 250 - 300 mL capacity?	((SM18)5210B,2.a (1992)		
2.2	Is the BOD incubator controlled at $20 \pm 1^{\circ}\text{C}$ and restrictive of exposure to light?	(SM18)5210B,2.b (1992)		
2.3	Is the dissolved oxygen measured with an electronic probe and meter that is calibrated before use and documented?	(SM18)4500-O,G,3.a(1992)		
2.4	Is the dissolved oxygen measured by iodometric titration (Winkler) using the azide modification method?	(SM18)4500-O,C,3.a(1992)		
2.5	Is the sodium sulfite solution prepared fresh daily?	(SM18)5210B,3.f (1992)		
2.6	Are reagents discarded when signs of biological growth or precipitation are observed?	(SM18)5210B,3.a (1992)		